

BIOCOMPATIBILITY CERTIFICATE

Testmaterial: Basys[®] Desinfect
Supplier: Basys Coating Systems
c/o Das Profil GmbH
Hohenfelderstrasse 17-19
D-22087 Hamburg

Studies performed: The following studies were performed in order to determine the biocompatibility of the product according to ISO 10993-1:

**CYTOTOXICITY
SENSITISATION (LLNA)
DERMAL IRRITATION**

Results: The product did not show any effect in the Sensitisation and Dermal Irritation study. A cytotoxic effect was obtained. Due to the intended use as a disinfectant, this effect could be anticipated.

BSL BIOSERVICE Scientific Laboratories GmbH Munich

Behringstraße 6/8

D-82152 Planegg



Dr. Daniela Stelter
Biological Safety Testing
Date: 12 May 2009



BASYS[®] ist sicher!

BASYS[®] COATING SYSTEMS GMBH